# Al for Experimental Controls at Jefferson Lab

Torri Jeske, Diana McSpadden, Naomi Jarvis, Kishan Rajput, Thomas Britton, and David Lawrence

AI4EIC



Thursday, September 2, 2021












## **Gluonic Excitations Experiment in Hall D**

 Designed to reconstruct exclusive final states from photoproduction reactions on proton targets


$$\vec{\gamma}p \rightarrow Xp$$

 Primary objective is to search for and measure exotic mesons





# **Calibrating Detectors is Time Consuming**




Time scale for calibrations is on the order of months



### Chamber gain calibration

- Calibration uses one raw data file (19GB) per run
  - -~450 production runs in 2018
  - -~950 production runs in 2020
- Gain Correction Factors are calculated based on most probable value of Landau fit to CDC pulse height

CDC amplitude, hits with drift time < 100ns (tracks, theta 28-32 deg, z 52-78cm)





#### **Time To Distance Calibration**

 Current calibration method produces 6 unique calibration constants from fit to data

$$d(t) = f_{\delta} \left( \frac{d_0(t)}{f_0} P + 1 - P \right)$$


$$f_{\delta} = a\sqrt{t} + bt + ct^3$$

$$f_0 = a_1\sqrt{t} + b_1t + c_1t^3$$

$$a = a_1 + a_2 |\delta|$$

$$b = b_1 + b_2 |\delta|$$

$$c = c_1 + c_2 |\delta|$$




Model development for ttod calibration constants is in very early stages

## Can we utilize AI to perform near real time calibrations?


Maintain consistent
detector response to
changing conditions via
recommending HV settings
for next run

Reduce time needed for calibrations

Make it adaptable to other detector systems




### **GCF** fluctuations with pressure




 GCFs are dependent on atmospheric pressure, temperature, current drawn by the high voltage boards, beam intensity, etc

#### Input features



- EPICS: Experimental Physics and Industrial Controls System
- Beam current, atmospheric pressure, thermocouple temperatures, pair spectrometer rates, high voltage board current, existing calibration constants
- Currently take data from one of multiple raw data files per production run

#### Gain calibration model architecture





#### **Sequential Neural Network**

- -122 total features
- Each feature is min-max scaled
- Loss function: % error between predicted GCF and existing GCF
- Optimizer: Adam, learning rate: 0.0002, with decrease learning rate on plateau used during training
- Early stopping if learning stopped for 50 epochs



### Model predictions and feature evaluation









## Very preliminary physics validation studies


- Existing calibration constants might not be the best truth to compare to
- Evaluate CDC resolution using predicted constants







## **Altered HV running conditions**




- HV settings for the next run may not be the same as the standard operating voltage!
- We don't want to recommend a HV setting that could trip the CDC or cause other problems

## **Adjusting CDC operating voltage**

- Current CDC operating voltage is 2125 V
  - This does not really change
  - Limited data obtained from HV scans
- For high and low gain scenarios, what is the associated high voltage?

#### CDC gain relative to that for standard HV






### **Estimating new HV running values**

Obtain new HV values associated with various GCFs at high/low pressure


Run	GCF	Pressure from EPICS (kPa)	Calibrated Pressure (mmHg)	GCF/GCF at std Pressure	New HV
51687	0.173	102.067	776	1.146	2137
51570	0.160	101.042	768	1.060	2129
51762	0.151	100.016	760	1.000	2125
51287	0.139	99.1262	753	0.921	2116
51160	0.132	98.4129	747	0.874	2111

New HV from fit to relative GCF as function of HV. Runs selected reflect the highest and lowest values of pressure seen during the 2018 run period.

## **Drift times with altered high voltage**



#### Garfield predictions for 50/50 Ar/CO2 and 1.8T





#### Recap + Ongoing work

#### Recap:

- Predict existing gain calibration constants with changing experimental conditions (2018 vs 2020 data)
- Established boundary for operating voltage of CDC based on previous run periods

#### Ongoing work:

- Evaluating CDC resolution with predicted calibration constants
- Time to Distance model development in progress
- Incorporating physics information into model
- Application to CLAS12 Drift Chambers located in Hall B

#### Contact information:

- Torri Jeske | roark@jlab.org
- Diana McSpadden | dianam@jlab.org



back ups



#### **GlueX Central Drift Chamber**

- Cylindrical, straw-tube wire chamber
  - 3522 straw tubes arranged in 28 layers:
    12 axial and 16 at offset from axial
  - Straws are 1.5 m long with 1.6 cm diameter
  - Gold-plated tungsten wire as the anode
  - -50:50 mix of Ar: flows through straws
- Used to detect and track charged particles with momenta p > 0.25 GeV/c
- Can also be used for PID using dE/dx





